



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. PHC, Melaramanathi

Lab No

12559

Constituent Group Chemistry III

Date of Result Entered :

20/05/2023

PT Item Lyophilized human serum based

Date of Report Published :

05/06/2023

Sl. No	Analyte	Reference Range	Analyzer Name	No. of Participants	DV	CV	Participating Lab CV	CV	SC	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	186 mg/dL	0.46	2.53		
2	CREATININE III	Jaffes Kinetic - Alkaline Picate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.7 mg/dL	-0.41	0.03		
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	114 mg/dL	-0.14	0.94		
SDI Range												
				Interpretation								
Within -1.00 to +1.00				Excellent.								
Within ±1.01 to ±2.00				Good.								
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.								
Beyond ±3.0				Unacceptable performance. Action Signal.								

LAB ADDRESS :

Govt. PHC, Melaramanathi
Melaramanathi
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****