



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. PHC, Melaramanathi

Lab No: 12559

Constituent Group: Chemistry III

Date of Result Entered: 19/06/2023

PT Item: Lyophilized human serum based

Date of Report Published: 04/07/2023

Sl. No.	Analyte	Reference / Probable Name	Analyzer Name	No. of Participants	CV	Participant 1	Participant 2	Participant 3	Participant 4	Participant 5	Participant 6	Participant 7	Participant 8	Participant 9	Participant 10
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	65	287.55	4.50	12.95	305 mg/dL	1.35	3.21					
2	CREATININE III	Jaffes Kinetic - Alkaline Pictate III	Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16	2.7 mg/dL	0.73	0.04					
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	136 mg/dL	4.61	0.76					
		SDI Range		Interpretation											
		Within -1.00 to +1.00		Excellent.											
		Within ±1.01 to ±2.00		Good.											
		Within ±2.01 to ±2.99		Accept with caution. Warning Signal.											
		Beyond ±3.0		Unacceptable performance. Action Signal.											

LAB ADDRESS :
 Govt. PHC, Melaramanathi
 Melaramanathi
 TAMILNADU

Coordinator Contact Details:
 Email: cllnqpc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****