



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Melaramanathi Lab No: 12559
 Constituent Group: Chemistry III Date of Result Entered: 17/07/2023
 PT Item: Lyophilized human serum based Date of Report Published: 07/08/2023

Sl. No.	Test Name	Method	Manufacturer	Lot No.	Expiry Date	CV	CV _{max}	CV _{min}	CV _{act}	CV _{act} %
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95	114	0.72	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Pictate III	Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27	6.7	2.99	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	127	2.91	0.80
SDI Range										
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							
Interpretation										

LAB ADDRESS :
 Govt. PHC, Melaramanathi
 Melaramanathi
 TAMILNADU

Coordinator Contact Details:
 Email: cdingc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paula Christudoss
Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****