

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 27.09.2023
		Calibration valid : 26.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVT.PRIMARY HEALTH CENTRE, KARNAMADAI BLOCK, CHINNAKALLIPATTI- 641302.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMOSCIENTIFIC
ii	Model	FINNIPIPETTE F3
iii	Value	10 µl -100 µl
iv	Serial No	RW01647

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com