

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :
 The Community Health Centre,
 Kannurpuzhar
 Virudhanganur-626204

CALIBRATION DATE: 23.05.2023
 EXPIRES ON DATE: 22.05.2024

CERTIFICATE NO: TICS-2023-KSI
 INSTRUMENT NAME: LABOTECH
 MANUFACTURE: LABOTECH
 SERIAL NO: 76663
 INSTALLED DEPARTMENT: LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL
DIGITAL THERMOMETER	BUZASHI 88110	RM2134.01	N104927

TEST REPORT

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE
ROTATION PER MINUTE	RPM	SPEED 2	1738	± 5
		SPEED 4	2300	± 5

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment available in National International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVE

Authorized Sign With Seal

*Insert serial when not used attached

CENTRIFUGE
 LABOTECH
 Model :- B01-153 (K123456)
 Serial no :- 76663 (A1049275)

Centrifuge Calibration on
 23/05/2023 / KSP/VMA/008
 Calibration no :- 23/05/2023
 Calibration Due :- 22/05/2024

(Signature)
 Digitally signed by
 K. S. Prasad
 DN: cn=K. S. Prasad, o=TICS