

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE AMATHUR, VIRUDHUNAGAR, TAMIL NADU-626001		CERTIFICATE NO	TICS/ALA/23-24/024
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE MODEL	THERMOSCIENTIFIC 100-1000 µL
CALIBRATION DATE	13/07/2023	SERIAL NO	Rw02603
CALIBRATION DUE DATE	12/07/2024	INSTALLED DEPARTMENT	LAB

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No


TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVC5PL22/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.2	± 1	PASS
		990	990.1	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Calibrated By:  Authorized Sign With Seal



TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED

No.763/19,BLOCK NO:10,KTC NAGAR,KULASEKARAPATTI,
TIRUNELVELI,TAMIL NADU, INDIA-627806.
Phone : 04633-210998 Mobile: +91 72000 04781, +91 95001 13759
E-Mail : transindiacalibration@gmail.com Website : www.transind.in
CIN: U52100TN2020PTC139842