

may-2023

## VIEW LAB MONTHLY SUMMARY


<b>Lab Name</b>	Govt. PHC, Periakulam	<b>Details About Robust Analysis</b>
<b>Lab No</b>	10628	<b>Detail About Monthly Summary</b>
<b>Month</b>	May	<b>Detail about SDI</b>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	
Click on the analyte to view Graphical Data		<input type="button" value="All Analyser Result"/> <input type="button" value="Print"/>
		Date of Result Entered : 15/05/2023
		Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	185 mg/dL	0.44	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	116.2 mg/dL	0.27	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

  
 Medical Officer,  
 Family Health Clinic,  
 Govt. Primary Health Centre,  
 Periyakulam T.V.Malal Dist