



VIEW LAB MONTHLY SUMMARY

Lab Name: GovL PHC, Perivakulam
 Lab No: 10028
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 08/06/2023

Date of Report Published : 04/07/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	BV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	3.17	14.82	303.97 mg/dL	1.57	0.69
2	CREATININE III	Jaffes Knetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.51 mg/dL	-0.35	0.91
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	3.07	117.21 mg/dL	0.69	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



Site Maintained by
Swiftware Solutions

Best Viewed at pixels 1024 X 768

[Signature]
 Medical Officer,
 Family Health Clinic,
 Govt. Primary Health Centre
 Perivakulam T.V.Malai Dist