



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Periyakulam	Details About Robust Analytic
Lab No	10629	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 06/07/2023

Date of Report Published : 07/08/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-PDD III	Robonik ( Semi automatic / Fully Automatic )	1976	109.35	3.55	6.07	114.9 mg/dL	0.91	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	3.83	0.00	0.35	3.89 mg/dL	0.17	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	112.1 mg/dL	0.45	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Logout