



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Periyakulam	<a href="#">Details About Robust Analysis</a>
Lab No	10628	<a href="#">Detail About Monthly Summary</a>
Month	August	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 10/09/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participant		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	173.59	3.13	3.03	173 mg/dL	-0.66	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1832	1.76	6.77	0.12	2.3 mg/dL	4.5	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.04	117 mg/dL	0.69	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

<a href="#">Home</a>	<a href="#">About Us</a>	<a href="#">Contact Us</a>	<a href="#">Feedback</a>
<a href="#">Links</a>	<a href="#">Privacy Policy</a>	<a href="#">Terms &amp; Conditions</a>	<a href="#">Disclaimer</a>
<a href="#">My Lab Address</a>	<a href="#">My My Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">My Summary</a>	<a href="#">Change Password</a>	<a href="#">Logout</a>	<a href="#">Sign-out</a>