

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE, AMATHUR, VIRUDHUNAGAR, TAMIL NADU-626001		CERTIFICATE NO	TICS/AMA/23-24/030
		EQUIPMENT NAME	REFRIGERATOR
		MANUFACTURE	WHIRLPOOL
		MODEL	ICEMAGIC
CALIBRATION DATE	13/07/2023	SERIAL NO	AMA-030
CALIBRATION DUE DATE	12/07/2024	INSTALLED DEPARTMENT	OPD

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DATA LOGGER WITH SENSOR	TEMPSONS	TPLR-96	2022-2023/eqn/2205005-01
			Temperature 25°C Humidity RH 55%

TEST REPORT

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
TEMPERATURE	°C	7	7.2	± 1	PASS
		6	6.3	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By: *Rer* Authorized Sign With Seal


TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED
 No.763/19,BLOCK NO:10,KTC NAGAR,KULASEKARAPATTI,
 TIRUNELVELI,TAMIL NADU, INDIA-627806.
 Phone : 04633-210998 Mobile: +91 72000 04781, +91 95001 13759
 E-Mail : transindiacalibration@gmail.com Website : www.transind.in
 CIN: U52100TN2020PTC139842