




CALIBRATION CERTIFICATE

CERTIFICATE NO.:	SBSICL2305207	MECHANICAL	Page No.1 of 1	
Issue Date:	01/04/2023			
SRF No & Date:	NR4/20190072-0001 & 18-05-2023			
Revised Date:	18-06-2023			
Calibration Date:	18-06-2023			
Calibration Due:	17-06-2024			
Customer Name & Address				
GOVERNMENT TAMILU I HEALTH CENTRE KUMARAVAKKAM-631001 CHEYSSARPATTI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description:	CENTRIFUGE	Make & Model:	RPM & R 80	
Range:	5000 RPM	SI No:	20441	
Resolution:	10 RPM	Identification No:	NA	
DUC Condition:	Reference	Location:	LABORATORY	
Environmental Conditions & Standard Operating Procedure Details				
Environmental Data:	Temperature: 24.3 °C	Relative Humidity:	34.5% RH	
Calibration Reference No:	SBSICRML04	Calibration done at:	ICNGCIL	
Reference Standards Details				
S.No	Description	Make & SI No:	Certificate No	
1	Legal Instrument	LINE GENI 175-0034V	1M407763-01	
			Valid to	
			10-11-2023	
CALIBRATION RESULTS				
S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (%)
	RPM	RPM	RPM	%
1	1000	999.5	0.2	4.2
2	2000	1999	0.1	4.2
3	3000	2999	0.2	4.2



SUNSHINE BIOMEDICAL SOLUTIONS
 No.68, First Floor, Padmasai Main Road, Kumbakonam, Chennai - 605002. Tel: 0436 - 2236 3967