



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, Madathupatti Lab No 12805  
Constituent Group Chemistry III Date of Result Entered : 18/08/2023  
PT Item Lyophilized human serum based Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	178.32 mg/dL	0.30	0.42
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	27.26 mg/dL	-0.24	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	2.05 mg/dL	2.44	0.01
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	149	113.96	4.36	4.97	113.6 mg/dL	-0.07	0.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. PHC, Madathupatti  
PAMBUKOIL SHANDHY  
Madathupatti  
TAMILNADU627862

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

[https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult\\_new.aspx?resDetId=268](https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDetId=268)

1/2

**MEDICAL OFFICER**  
Government Primary Health Centre  
Madathupatti - 627 862