



CALIBRATION CERTIFICATE

MEDICAL DEVICES		Page No:1 of 1		
CERTIFICATE NO: SBS/CL/23/10602				
Issue Date	29-08-2023			
SRF No & Date	SRF/23/00338-0002 & 28-08-2023			
Receipt Date	28-08-2023			
Calibration Date	28-08-2023			
Calibration Due	27-08-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, LALIGAM-636804, DHARMAPURI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	ELECTRICAL SAFETY (CELL COUNTER)	Make & Model : SYSMEX & XP 300		
Range	MULTI	Sr. No : B5542		
Resolution	MULTI	Identification No : NA		
DUC Condition	SATISFACTORY	Location : LABORATORY		
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature: 25.6 °C	Relative Humidity : 52% RH		
Calibration Procedure No	SBS/CP/MD/29	Calibration done at : ONSITE		
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-202


ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	97	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μAfor B,BF,CF	175	7.5
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μAfor B,BF,CF	228	7.2


REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 C.SIVABALAN



Authorised Signatory

 Technical Manager
 C.SHANMUGARAJ

Chief Executive