



## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12101		MEDICAL DEVICES		Page No:1 of 1
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00479-0002 & 29-09-2023			
Receipt Date	29-09-2023			
Calibration Date	29-09-2023			
Calibration Due	28-09-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT COMMUNITY HEALTH CENTRE, THIRUPPUKUZHI-631551,KANCHIPURAM DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	: SEMI AUTO ANALYSER	Make & Model	: ROBONIK & PRIETEST TOUCH	
Range	: MULTI	Sr. No	: ATCD0390519RBK	
Resolution	: MULTI	Identification No	: NA	
DUC Condition	: SATISFACTORY	Location	: LABORATORY	
<b>Environmental Conditions &amp; Calibration Procedure Details</b>				
Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

#### RESULTS


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 87	Uncertainty in % (±) 13.92
2	Earth Leakage <5000μAfor B,BF,CF	Measured values in μA 195	Uncertainty in % (±) 9.3
3	Enclosure Leakage <500μAfor B,BF,CF	Measured values in μA 241	Uncertainty in % (±) 8.4

#### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
M.ADHIBAN



  
Technical Manager  
(C.SHANMUGARAJ)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.