

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/11769 MEDICAL DEVICES Page No:1 of 1

Issue Date	26-09-2023
SRF No & Date	SRF/23/00402-0006 & 25-09-2023
Receipt Date	25-09-2023
Calibration Date	25-09-2023
Calibration Due	24-09-2024

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
VELLIRAVELI-641603.TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : ATCD2960321R8K
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LAB

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 88	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 169	Uncertainty in % (±) 7.6
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 229	Uncertainty in % (±) 7.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



C. Shanmugaraj
Technical Manager
C.SHANMUGARAJ

Chief Executive