

< 566f20be-5772-49b0-9d2b-f932d7c3d944



SARAVANA MEDICO

Shop No. 41/50, Good Shed Road, 4th Street, Adambakkam, Chennai - 600 088
 E-mail : saravanamedico2019@gmail.com, Mobile : 7358077155

CALIBRATION CERTIFICATE		Page No	01 of 02		
CERTIFICATE NO: LDM/CAL/KPHC/0323/SA01		Temp	: 25±4 °C		
NAME AND ADDRESS OF THE CUSTOMER		Humidity	: 30-75 RH%		
The Medical Superintendent, Government PHC, Kundrathur, Chennai - 600069		Date of Issue	: 16.03.2023		
ULR.NO: ---		Date of Receipt of item	: 16.03.2023		
Date of Calibration		Date of Calibration	: 16.03.2023		
Suggested date of Calibration		Suggested date of Calibration	: 15.03.2024		
Ref. Documents		Ref. Documents	: As per Manufacturer specification		
Cal. Procedure Ref. No.		Cal. Procedure Ref. No.	: LDM-CP-MD19		
Customer Ref No.		Customer Ref No.	: Verbal		
DESCRIPTION OF DUC					
Name of the DUC	: Semi Auto Analyser	Receiving Condition of the DUC	: Good		
Make	: ---	Location	: Laboratory		
Model	: ---	Calibrated at	: Onsite		
Sl.no.	: ---	Range	: Multi		
ID No.	: ---	Resolution	: Multi		
DETAILS OF MASTER USED					
Standard Instrument Used	Make / Model	ID No./ Sl.No.	Traceable to / Certificate No.	Validity	
Electrical Safety Analyzer	Datrend /vPad Rugged 2	VRR20070452	2780622-E01	22/6/2023	
The Master are traceable to National / International standards.					
VISUAL INSPECTION				Remarks	
SL.No.	Description			Pass	Fail
1	Device Clean and Decontaminated			✓	
2	Physical damage			✓	
3	Power Cord check			✓	



Am...
 PAN...
 GSTIN...
 State...
 Doc...
 No...
 ar...
 Tern...
 nec...
 be...
 yn...