



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-10

Lab Name: *CLINICAL BIOCHEMISTRY* Lab No: *12345*  
 Constituent Group: *CLINICAL* Date of Result Entered: *18/08/2023*  
 PT Item: *Lyophilized human serum based* Date of Report Published: *21/08/2023*

Sl. No.	Test Name	Method	Instrument	Result 1	Result 2	Result 3	Result 4	Reference Range	SDI
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	987 mg/dL	1.27
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.43 mg/dL	-1.09
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	112 mg/dL	-0.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### LAB ADDRESS :

*CLINICAL BIOCHEMISTRY*  
 CHRISTIAN MEDICAL COLLEGE, VELLORE  
 TAMIL NADU

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*