

LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Anna Nagar  
 Lab No 10927  
 Month June  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 21/06/2023

Date of Report Published : 04/07/2023

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42	270 mg/dL	-0.92	3.81	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17	2.6 mg/dL	-0.40	0.06	
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81	115 mg/dL	0.25	1.58	

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

*[Signature]*  
 Medical Officer  
 Anna Nagar UPHC  
 Tiruvannamalai-606603