

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Anna Nagar
 Lab No 10927
 Month February
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 21/02/2023

Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	153.37	8.21	12.59	137 mg/dL	-1.30	4.26
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	37	3.35	7.70	0.26	3.3 mg/dL	-0.19	0.08
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	113.39	6.52	7.39	107 mg/dL	-0.86	2.57

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


Medical Officer
Anna Nagar UPHC
Tiruvannamalai-606603.

Links

- [Modify Address](#)
- [Modify Methods](#)
- [Submit Lab Results](#)
- [View Lab Monthly Summary](#)

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Anna Nagar	Details About Robust Analysis
Lab No	10927	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 21/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84	93 mg/dL	0.04	2.61
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13	1.3 mg/dL	-0.54	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68	118 mg/dL	0.88	2.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


Medical Officer
Anna Nagar UPHC
Tiruvannamalai-606603.

Links

- [Modify Address](#)
- [Modify Methods](#)
- [Submit Lab Results](#)
- [View Lab Monthly Summary](#)

[View Monthly Summary](#) [View Result 2](#) new.aspx?resDetId=263&noSamples=1&month=3&year=2023&groupId=9&yearGroup=110

VIEW LAB MONTHLY SUMMARY

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Lab Name Govt. UPHC, Anna Nagar

Lab No 10927

Month April

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 21/04/2023 ✓

Date of Report Published : 06/05/2023 ✓

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67	129 mg/dL	1.10	2.03
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	2.5 mg/dL	2.19	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	121 mg/dL	1.37	2.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Noted
 20/05/2023
 Dr. N. S. S. (S-1)

Medical Officer
 Anna Nagar UPHC
 Tiruvannamalai-606603

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Anna Nagar
 Lab No 10927
 Month May
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

All Analyser Result

Date of Result Entered : 22/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85	179 mg/dL	-0.08	3.4
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	46	29.96	11.88	3.56	32 mg/dL	0.57	1.0
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16	1.6 mg/dL	-1.13	0.0
4	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	113 mg/dL	-0.47	2.4

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

[Signature]
 Medical Officer
 Anna Nagar UPHC
 Tiruvannamalai-606603.

Links

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=265&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110

LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Anna Nagar
 Lab No: 10927
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 21/06/2023

Date of Report Published : 04/07/2023

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42	270 mg/dL	-0.92	3.81	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17	2.6 mg/dL	-0.40	0.06	
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81	115 mg/dL	0.25	1.58	

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

[Signature]
 Medical Officer
 Anna Nagar UPHC
 Tiruvannamalai-606603