

2 & 3, Devi Illam, Priya Nagar, Telungupalayam Pirivu, Coimbatore - 641 039, TamilNadu.

GSTIN: 33APLPA1828A1ZM

T. ARUN PRASAD, M.Tech., MBA., (Ph.D)

Managing Director

Consultant Biomedical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : MICROSCOPE

Protect Class/Type:

Risk class II/CF

Make

: BINACULAR

S.No

3.20009E+11

Model

: MLW

Location

MELPALLIPATTU - GPHC

S.no		Measuremen	t Parameters	Status	Remarks
1	Electrical Test	(b). Equipo (c). Risk of (d). Chassis (e). Earth &	Standards (range: 210-250 V) Itential ground &Earthing Shock/fire Hazards, if any S&Insulation Resistance & Chassis Leakage Current Condition & charging status	235volts 0 to 5volts No Zero ohms 10mA	Stabilizer/UPS With in ranges well
11	Physical Condition Test	(b). Power Power	 (a). Cabinet & Visible damages, if any (b). Power supply unit, halogen lamps, Power cord accessories (c). Risk of mechanical damages, if any 		Checked ok
111	Performance Test	(a). haloge (b). intens a parti (c). Measu	 (a). halogen lamps intensity (b). intensity test taken by lux meter for a particular lamp & results (c). Measurement of Lux/Lens range (d). Lamp life time 		Fine brightness perfect 20000 hrs appx
IV	Calibration Test	(a). LED la (b). intens a part (c). Measu	(a). LED lamps intensity (b). intensity test taken by lux meter for a particular lamp & results (c). Measurement of Lux/Lenses cal (d). Lamp life time		In range peak in range in range
Formulated By: Einstein Research Laboratory			Issued By:	User Department:	
Endorsed By: Research team head		Tap	98(1 149 - HAOTABINIO)	GJES on pany	
	wed On : 25.04.2 3 Review : 25.10.23		Kuedruga Golfensur Rediperur Cell. 98652.857.54. 8124517	2050.	