



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 12.09.2023
	Calibration valid : 11.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE , VADAKKANKULAM, THIRUNELVELI
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	CENTRIFUGE
i Manufactured by	LABOTECH
ii Model	BD-152
iii Serial No	PHC/VK/CENT/001

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	0	1.12
2	1200	1202	2	1.23
3	1500	1501	1	1.56
4	2000	2006	6	1.175

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016



GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662



No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com