

Lab Name Govt. UPHC, Anna Nagar

Lab No 10927

Month August

Year 2023

Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

[Click on the analyte to view Graphical Data](#)

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48	150 mg/dL	-2.32
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	1.5 mg/dL	-2.28
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49	113 mg/dL	-0.44
SDI Range Interpretation									
Within -1.00 to +1.00 Excellent.									
Within ±1.01 to ±2.00 Good.									
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.									
Beyond ±3.0 Unacceptable performance. Action Signal.									

Links

Modify Address Type here to search

Modify Methods     

Submit Lab Results

[View Lab Monthly Summary](#)

ENG 11:14 AM 9/30/2023