

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Labbaikudikadu

[Details About Robust Analysis](#)

Lab No 7126

[Detail About Monthly Summary](#)

Month July

[Detail about SDI](#)

Year 2023

Constituent Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

SL No	Analyte	Method / Principle Name	Analysed Name	No of Participants	DV	CV	SD	Year Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	121 mg/dL	1.92	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.2 mg/dL	1.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	113 mg/dL	0.00	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Verify Summary	Change Password	Home	Sign-out



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. Block PHC, Labbaikudikadu
 Lab No: 7126
 Constituent Group: Chemistry III
 Date of Result Entered: 19/06/2023
 PT Item: Lyophilized human serum based
 Date of Report Published: 04/07/2023

2023.09.30 11:40

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	268 mg/dL	-1.26	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.55 mg/dL	-0.12	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	115 mg/dL	0.25	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within \$1.01 to \$2.00				Good.						
Within \$2.01 to \$2.99				Accept with caution. Warning Signal.						
Beyond \$3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Labbaikudikadu
 The Block Medical Officer, Labbaikudikadu, Labbaikudikadu Po, Kunnan Tk, Perambalur Dt.
 Labbaikudikadu
 TAMILNADU621108

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paula Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. Block PHC, Labbaikudikadu
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based
 Lab No: 7128
 Date of Result Entered: 22/05/2023
 Date of Report Published: 05/06/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	172 mg/dL	-0.95	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	117 mg/dL	0.41	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Labbaikudikadu
 The Block Medical Officer, Labbaikudikadu, Labbaikudikadu Po, Kunnam Tk, Perambalur Dt.
 Labbaikudikadu
 TAMILNADU621108

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2233102
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

Pamela Christudoss



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023



2023.09.30 11:40

Lab Name: Govt. Block PHC, Labbaikudikadu

Lab No: 7126

Constituent Group: Chemistry III

Date of Result Entered: 20/04/2023

PT Item: Lyophilized human serum based

Date of Report Published: 06/05/2023

7126
 20/04/2023
 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Year Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	113 mg/dL	-0.39	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	114 mg/dL	0.41	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Labbaikudikadu
 The Block Medical Officer, Labbaikudikadu, Labbaikudikadu Po, Kurnam Tk, Perambalur Dt.
 Labbaikudikadu
 TAMILNADU621108

Pamela Christodoss

Coordinator Contact Details:
 Email:clnqg@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name: Govt. Block PHC, Labbakudikadu
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 7126
 Date of Result Entered: 20/03/2023
 Date of Report Published: 04/04/2023

Item No	Test Name	Method	Result	Reference Range	Units	Delta	Delta %
1	GLUCOSE III	GOD-POD III	1878	91.48	7.46	6.82	91
							mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1851	1.33	9.08	0.12	11
							mg/dL
3	CHOLESTROL III	CHOD-PAP III	1761	110.93	5.23	5.80	115
							mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance, Action Signal.

LAB ADDRESS :
 Govt. Block PHC, Labbakudikadu
 The Block Medical Officer, Labbakudikadu, Labbakudikadu Po, Kunnam TK, Perambalur Dt.
 Labbakudikadu
 TAMILNADU621108

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****