

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Thirumuruganpoondi  
 Lab No: 10621  
 Month: May  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

All Analyser Result

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

82  
 May '23

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD	SDI			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	195 mg/dL	1.50	0.44	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.79 mg/dL	0.23	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	114.1 mg/dL	-0.09	0.27	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

*M. Sujal*  
**MEDICAL OFFICER**  
 URBAN PRIMARY HEALTH CENTER  
 THIRUMURUGANPOONDI - 641 952

Links

[https://home.cmcvellore.ac.in/clinql/ViewMonthlySummary/view/Result2\\_new.aspx?resDetId=265&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110](https://home.cmcvellore.ac.in/clinql/ViewMonthlySummary/view/Result2_new.aspx?resDetId=265&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110)