



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023

6
March/23



PC-1024

Lab Name	Govt. PHC, Thirumuruganpoondi	Lab No	10621
Constituent Group	Chemistry III	Date of Result Entered :	18/03/2023
PT item	Lyophilized human serum based	Date of Report Published :	04/04/2023

S. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SD	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	87.6 mg/dL	-0.57	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.45 mg/dL	0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	115.4 mg/dL	0.77	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Thirumuruganpoondi
THRUMURUGAN POONDI TIRUPUR District ..
Thirumuruganpoondi
TAMILNADU 641652

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2233102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
URBAN FAMILY HEALTH CENTER
THRUMURUGAN POONDI - 641 652