

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Idikarai	Details About Robust Analysis								
Lab No	12036	Detail About Monthly Summary								
Month	March	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data		All Analyser Result	Print							
Date of Result Entered : 21/03/2023										
Date of Report Published : 04/04/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	88.5 mg/dL	-0.44	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.1 mg/dL	-1.90	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	104.4 mg/dL	-1.13	0.28
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

L. Singh
 MEDICAL OFFICER
 GOVT PRIMARY HEALTH CENTRE
 IDIKARAI, COIMBATORE (DIST)