

6/20/23 12:11 PM



External Quality Assurance Scheme - Post Monthly Summary

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name: Govt. Block PHC, Kaniyambadi
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7660
Date of Result Entered: 20/03/2023
Date of Report Published: 04/04/2023

SI No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	Score Value	SD	Z
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.48	6.82	85 mg/dL	-0.95	0.21
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.81
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	108 mg/dL	-0.50	0.28
		SDI Range								
		Within -1.00 to +1.00								
		Within ±1.01 to ±2.00								
		Within ±2.01 to ±2.99								
		Beyond ±3.0								
		Interpretation								
		Excellent.								
		Good.								
		Accept with caution. Warning Signal.								
		Unacceptable performance. Action Signal.								

LAB ADDRESS :
Govt. Block PHC, Kaniyambadi
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VELLORE DISTRICT,
Kaniyambadi,
TAMILNADU632102

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Chelvanathan
Dr. Pamela Chelvanathan
CMC EDAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EDAS reports is confidential