



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. Block PHC, Kaniyambadi  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 7666  
Date of Result Entered: 19/04/2023  
Date of Report Published: 06/05/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	SD	Year Value	SD
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	126 mg/dL	0.08
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.4 mg/dL	1.37
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	121 mg/dL	1.62

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. Block PHC, Kaniyambadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VELLORE DISTRICT.  
Kaniyambadi  
TAMILNADU632102

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential