



# G MULTI SERVICES



78, Yahooob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin:638 452.  
gmultiservicesgobi@gmail.com. Cell : 96597 20654.  
Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE., BE (ICE)  
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : CELL COUNTER  
MAKE : SYSMEX  
MODEL : XP-100

PROTECT CLASS /TYPE : Risk Class I  
S. No. : B5886  
Location : UGPHC, KANNIYAMBADI

S.No	Measurement Parameters	Status	Remarks
1	<b>Electrical Test</b> (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS  With in Ranges Well
2	<b>Physical Condition Test</b> (a) Cabinet & Visible damages, if any (b) Front Panet/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK No	Checked OK
3	<b>Performance Test</b> (a) On /Off switch (b) switch & functional tests (c) Spindle Motor (d) Running test (e) Abnormalities (f) Display checkup (g) Keypad/indicators & connector	Normal OK OK OK No Well Well	Ok Perfect Perfect
4	<b>Calibration Test</b> (a) WB Test (b) Printer (c) Display checkup (d) Timer test (e) Result	Done Ok OK OK Perfect	Perfect
<b>Formulated By:</b> Edition Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>
<b>Endorsed By :</b> Research team head 			
Reviewed On: 15.08.23 Next Review : 14.08.24			