



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



10539

20/06/2023

04/07/2023

Lab Name Govt. PHC, Aralvaimozhi  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No  
Date of Result Entered :  
Date of Report Published :

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	
						CV	SD		SDI	✓
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	279.5 mg/dL	-0.48	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	118.2 mg/dL	0.88	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent. ✓
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. PHC, Aralvaimozhi  
CLINICAL LABORATORY, KANYAKUMARI(DIST)  
Aralvaimozhi  
TAMILNADU629301

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

யமுதலய அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
ஆரலவாய்மொழி - 629 301  
தமிழ்நாடு சுகாதார அமைச்சு