



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



10539

Lab Name Govt. PHC, Aralvaimozhi
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No
Date of Result Entered : 20/06/2023
Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	279.5 mg/dL	-0.48	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	118.2 mg/dL	0.88	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent. ✓
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Aralvaimozhi
CLINICAL LABORATORY, KANYAKUMARI(DIST)
Aralvaimozhi
TAMILNADU629301

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

யமுதலய அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
ஆரவமாய்மொழி - 629 301
தமிழ்நாடு சுகாதார அமைச்சு