



CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVERNMENT UPGRADE PRIMARY HEALTH CENTRE PAVOORCHATHRAM, TAMIL NADU-627806		CERTIFICATE NO	TICS/PH/2023/01/11/062
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE	MICROLUX
		MODEL	10-100 µL
CALIBRATION DATE	02/01/2023	SERIAL NO	Ph-062
CALIBRATION DUE DATE	01/01/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPL22/07/884-01	27-JUL-2023

TEST REPORT

					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.2	± 1	PASS
		99	99.1	± 1	PASS

* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traccability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE
 Calibrated By: _____ Authorized Sign With Seal



No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.

☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759

✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com

CIN: U52100TN2020PTC139842



CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS : GOVERNMENT UPGRADE PRIMARY HEALTH CENTRE PAVOORCHATHRAM, TAMIL NADU-627806		CERTIFICATE NO	TICS/PH/2023/01/11/065
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE	MICROLUX
		MODEL	5-50 µL
CALIBRATION DATE	02/01/2023	SERIAL NO	PH-065
CALIBRATION DUE DATE	01/01/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP122/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.1	± 1	PASS
		49	49.1	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Calibrated By: Authorized Sign With Seal

[Signature]



[Signature]



CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVERNMENT UPGRADE PRIMARY HEALTH CENTRE PAVOORCHATHRAM, TAMIL NADU-627806		CERTIFICATE NO	TICS/PH/2023/01/11/063
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE MODEL	THERMOSCIENTIFIC 100-1000 µL
CALIBRATION DATE	02/01/2023	SERIAL NO	RW12043
CALIBRATION DUE DATE	01/01/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPL22/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.2	± 1	PASS
		999	999.3	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE
 Calibrated By: _____ Authorized Sign With Seal





CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVERNMENT UPGRADE PRIMARY HEALTH CENTRE PAYOORCHATHRAM, TAMIL NADU-627806		CERTIFICATE NO	TICS/PH/2023/01/11/061
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE	THERMO SCIENTIFIC
		MODEL	10-100 µL
CALIBRATION DATE	02/01/2023	SERIAL NO	RW09397
CALIBRATION DUE DATE	01/01/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP22/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.5	± 1	PASS
		99	99.6	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Calibrated By: _____ Authorized Sign With Seal

[Signature]



[Signature]



CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :			CERTIFICATE NO	TICS/PH/2023/01/11/064	
GOVERNMENT UPGRADE PRIMARY HEALTH CENTRE			EQUIPMENT NAME	MICROPIPETTE	
PAVOORCHATHRAM,			MANUFACTURE	MICROLUX	
TAMIL NADU-627806			MODEL	100-1000 µL	
CALIBRATION DATE	02/01/2023	SERIAL NO	PH-064		
CALIBRATION DUE DATE	01/01/2024	INSTALLED DEPARTMENT	LAB		
<p>1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP122/07/884-01	27-JUL-2023		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.1	± 1	PASS
		999	999.1	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		