



Trust & Care



heartbeat

Medical Systems

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 12.09.2023
		Calibration Valid: 11.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALRH CENTRE, THULUKKARAPATTI, THIRUNELVELL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BDI 152
iii	Serial No	PHC/THI/CEN/001
iv	Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1509	9
4	2000	2000	0

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navalindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com