

# THE PIPETEMANN

An ISO 17025:2005 CERTIFIED NABL ACCREDITED MICROPIPETTE CALIBRATION LABORATORY  
No. 22, First Floor, Sri Sairam Bhavan, Prashanthi Nagar, K.R. Puram, Avarampalayam Road, Coimbatore - 641 006,  
Tamil Nadu, India. Ph. : 0422 - 2568182 / 6452043 Mobile : 98430 59043  
E-mail : customercare@pipettemann.in Website : www.pipettemann.in



## CALIBRATION CERTIFICATE

CUSTOMER ADDRESS :  
M/s. GOVT PRIMARY HEALTH CENTRE(TPM 364)  
ARALVAIMOZHY

KANYAKUMARI DIST 0  
Phone No. : 9444843948

NAME OF EQUIPMENT : MICROPIPETTE

CERTIFICATE NO. : TPMCC13109757

DATE OF ISSUE : 04/10/2013

RECEIPT NO. : 1956

TEMPERATURE : 20.7 °C

RECEIPT DATE : 03/10/2013

HUMIDITY : 50 %

JOB CARD NO. : TPM364101307294

AIR PRESSURE : 963 hpa

BRAND : ERBABIOHIT

MEDIUM OF WEIGHING : GRADE III WATER

RANGE : 100 - 1000 µl

DENSITY CORRECTION FACTOR [Z]  
'z' VALUE : 1.0031 µl/mg

SERIAL NO. : 6082812

TEST METHOD : ISO 8655 PART - 6

TIPS USED : 1000 µl

WEIGHING BALANCE : SEMI MICROBALANCE

SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE ( 1 µl - 10000 µl )

### TRACEABILITY ON CALIBRATION OF WEIGHING BALANCE

MAKE : SHIMADZU

SERIAL NO. : D450020184

RESOLUTION : 0.00001g upto 82g & above is 0.0001g

CALIBRATED BY : FCRI, PALAKKAD

CERTIFICATE NUMBER : PSL/718/0713/M/642

CALIBRATED DATE : 11/07/2013