

# THE PIPETEMANN

An ISO 17025:2005 CERTIFIED NABL ACCREDITED MICROPIPETTE CALIBRATION LABORATORY  
No. 22, First Floor, Sri Sairam Bhavan, Prashanthi Nagar, K.R. Puram, Avarampalayam Road, Coimbatore - 641 006,  
Tamil Nadu, India. Ph. : 0422 - 2568182 / 6452043 Mobile : 98430 59043  
E-mail : customercare@pipetemann.in Website : www.pipetemann.in



CERT. NO. C - 0755

## CALIBRATION CERTIFICATE

CUSTOMER ADDRESS :  
M/s. GOVT PRIMARY HEALTH CENTRE(TPM 364)  
ARALVAIMOZHY

KANYAKUMARI DIST 0  
Phone No.: 9444843948

NAME OF EQUIPMENT : MICROPIPETTE

CERTIFICATE NO. : TPMCC13109756

DATE OF ISSUE : 04/10/2013

RECEIPT NO. : 1956

TEMPERATURE : 21.9 °C

RECEIPT DATE : 03/10/2013

HUMIDITY : 56 %

JOB CARD NO. : TPM364101307293

AIR PRESSURE : 964 hpa

BRAND : VERTEX

MEDIUM OF  
WEIGHING : GRADE III WATER

RANGE : 5 - 50 µl

DENSITY CORRECTION FACTOR [Z]  
'z' VALUE : 1.0033 µl/mg

SERIAL NO. : NIL

TEST METHOD : ISO 8655 PART - 6

TIPS USED : 200 µl

WEIGHING  
BALANCE : MICROBALANCE

SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE ( 1µl - 10000 µl )

### TRACEABILITY ON CALIBRATION OF WEIGHING BALANCE

MAKE : SARTORIUS MICRO BALANCE

SERIAL NO. : 28001417

RESOLUTION : 0.000001g upto 6.1g

CALIBRATED BY : FCRI, PALAKKAD

CERTIFICATE NUMBER : PSL/2638/0313/M/2029

CALIBRATED DATE : 26/03/2013