

CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/01643	MECHANICAL	Page No:1 of 1
Issue Date	08-02-2023		
SRF No & Date	SRF/23/00050-0012 & 08-02-2023		
Receipt Date	08-02-2023		
Calibration Date	08-02-2023		
Calibration Due	07-02-2024		

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
MELARAMANATHI-623603, RAMANATHAPURAM DISTRICT.

Details of Device Under Calibration (DUC)

Description	CENTRIFUGE	Make & Model	REMI
Range	3500 RPM	Sr. No	
Least Count	10 RPM	Identification No	
DUC Condition	Satisfactory	Location	LAB

Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 25.4 °C	Relative Humidity	54% Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	TMS/22/63 01	10-11-2023

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	500	499.2	0.8	4.2
2	2000	1998.4	1.6	4.2
3	3000	2998.2	1.8	4.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

S. Murali
(Calibration Engineer)
S.MURALI



Authorised Signatory

Quality Manager Chief Executive
 (D.VETRI SELVI)