



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
All Analyser Summary Report - March 2023



Name
Instituent Group
Item

Govt. PHC, Begarahalli
Chemistry III
Lyophilized human serum based

Lab No
Date of Result Entered
Date of Report Published

10520
18/03/2023
04/04/2023

No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84		
			Transasia / Erba	18	92.24	11.77	10.85		
			Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	88 mg/dL	-0.51
			Any Analyser (Automation / Semi Automation)	74	92.52	7.71	7.13		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13		
			Transasia / Erba	16	1.36	9.29	0.13		
			Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.28 mg/dL	-0.41
			Any Analyser (Automation / Semi Automation)	70	1.32	7.52	0.10		
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68		
			Transasia / Erba	19	111.05	6.05	6.72		
			Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	108.23 mg/dL	-0.47
			Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78		

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within +1.01 to +2.00	Good.
within +2.01 to +2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

Govt. PHC, Begarahalli
LABORATORY, BEGARAHALLI PHC, KARIMANGALAM, DHARMAPURI DIST
BEGARAHALLI
TAMILNADU, 636906

Coordinator Contact details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not subcontract any components
***** End of Report *****

Arasu
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
பேகார அள்ளி-636 806.



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
All Analyser Summary Report - April 2023



Lab Name
Constituent Group
PT item

Govt. PHC, Begarahalli
Chemistry III
Lyophilized human serum based

Lab No
Date of Result Entered
Date of Report Published

10520
20/04/2023
06/05/2023

Sl No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67			2.93
			Transasia / Erba	19	118.42	8.48	10.04			4.61
			Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119.92 mg/dL	0.07	0.34
			Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.83			1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15			0.05
			Transason / Erba	19	2.14	6.03	0.13			0.06
			Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.25 mg/dL	0.51	0.01
			Any Analyser (Automation / Semi Automation)	65	2.14	7.02	0.15			0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03			2.07
			Transasia / Erba	20	111.15	4.25	4.72			2.11
			Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	108.13 mg/dL	-0.60	0.28
			Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79			0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

Govt. PHC, Begarahalli
THE MEDICAL OFFICER, CLINICAL LABORATORY, BEGARAHALLI PHC, KARIMANGALAM, DHARMAPURI DIST
Begarahalli
TAMILNADU, 636906

Coordinator Contact details:
Email: clinq@cmcvellore.ac.in
Contact Number: 0116-2283102

Pamela Christaloss
Dr. Pamela Christaloss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

M. J. S. S.
9/5/23
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
பேகராசுனி-636 809.

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. PHC, Begarahalli
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 19928
 Date of Result Entered: 18/05/2023
 Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SO			
1	GLUCOSE III	GCO-POD III	Roche (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	144 mg/dL	0.23	0.44
2	CREATININE III	Jaffe Kinetic - Alkaline Picrate	Roche (Semi automatic / Fully Automatic)	1854	1.79	7.48	0.13	1.79 mg/dL	0.23	0.81
3	CHOLESTROL III	CHOD-PAP III	Roche (Semi automatic / Fully Automatic)	1798	114.00	5.08	5.53	198 mg/dL	-1.13	0.77

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±3.00	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:
 Govt. PHC, Begarahalli
 THE MEDICAL OFFICER, CLINICAL LABORATORY, BEGARAHALLI PHC, KARIMANGALAM DHARMAPURI DIST
 Begarahalli
 TAMILNADU 638906

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 9449-2281102

Pamela Christudas
 Dr. Pamela Christudas
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

M. Jeyapalan
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 ஓகார அக்சன்-635 008.

Dr. Jeyapalan
 District Microbiologist



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



19528

Lab Name: Govt. PHC, Begarahalli
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 1898/2023
 Date of Result Entered: 01/09/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SCI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	171.58	5.13	9.08	177.46 mg/dL	6.21
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.74	6.77	0.12	1.68 mg/dL	-6.67
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1784	113.80	4.67	4.44	113.88 mg/dL	0.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Begarahalli
 THE MEDICAL OFFICER, CLINICAL LABORATORY, BEGARAHALLI PHC, KARIMANGALAM, DHARMAPURU DIST
 Begarahalli
 TAMILNADU636904

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0419-2293102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 — End of Report —

M. S. Srinivasan
 மருத்துவ அலுவலர்
 அரசு ஆரவ் சுகாதார பி...
 பெகராஹ்லி

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Begarhalli
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 10520
 Date of Result Entered: 19/07/2023
 Date of Report Published: 07/08/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants			Your Value	SD
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	109.35	15.95	8.97	118 mg/dL	1.43	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1950	5.83	8.00	8.29	8.78 mg/dL	4.14	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1778	113.99	4.12	4.96	97.23 mg/dL	3.38	

SD Range	Interpretation
Within +1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Begarhalli
 THE MEDICAL OFFICER, CLINICAL LABORATORY, BEGARAHALLI PHC, KARIMANGALAM, DHARMAPURI DIST
 Begarhalli
 TAMILNADU 636906

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Signature]
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 பேகரா அள்ளி-636 906