Thrissur Office:

Thattil Nadakalan Complex, Kuriachira P.O.,

Thrissur - 680 006, INDIA.

Tel.: (0487) 2250793, 6544048. Mobile No.: 9645084048.

E-mail: tulipkerala@gmail.com

27.09.2023

CIN: U85195GA1988PTC005717

## **CALIBRATION CERTIFICATE**

This is to certify that, CORALAB TOUCH Semi-Automatic Analyzer bearing Serial No BS4GS046E

Installed at, THE MEDICAL OFFICER, GOVT.UPHC, NANJUNDAPURAM, COIMBATORE is performing satisfactorily.

CORALAB TOUCH is calibrated last on 27.09.2023 by our Service Engineer,

Mr. JEBARAJAN

Following checks have been performed:

- 1. Peristaltic Pump Cleaned
- 2. Cleaned flowcell & Tubings
- 3. Adjusted Lamp Gain.
- 4. Checked all the Voltage.

Next Calibration Due on 26.09.2024

For TULIP

AUTHORISED SIGNATORY

Corporate Office: Gitanjali, Tulip Block, Dr. Antonio Do Rego Bagh, Alto Santacruz, Bambolim Complex P.O., Goa - 403 202, INDIA. Tel.: (0832) 2458546-51, Fax: (0832) 2458544. E-mail: sales@tulipgroup.com, Website: www.tulipgroup.com