

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Puluvaipatti
 Lab No 7343
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

No.	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	Mean	SDI	CV	View
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	176 mg/dL	0.05	0.42	
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	25 mg/dL	-0.77	0.65	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2 mg/dL	2.02	0.01	
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Mr. Hanthamad
04/09/23.

மருத்துவ அலுவலர்
 சிம்கி கிளப்பு கார்டர் நியமிக்கப்பட்டவர்
 மருத்துவ அலுவலர்
 கோவைத்திட்டம் - 641 101