

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Puluvapatti
 Lab No 7343
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	176 mg/dL	0.05	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	25 mg/dL	-0.77	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2 mg/dL	2.02	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Quality Address	Modify Methods	Submit Lab Results	View Lab Monthly
Weekly Summary	Change Password	Home	Sign-out

Mr. Hanthamad
04/09/23

மருத்துவ அலுவலர்
 சுக. அலுவலகம்
 திருச்சி, தமிழ்நாடு - 621 001