



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, Pethampalayam Lab No 12199
Constituent Group Chemistry III Date of Result Entered : 19/08/2023
PT item Lyophilized human serum based Date of Report Published : 01/09/2023

Sl.No	Analyte Name	No of Participants	Participants DV	Participants		Your Value	SDI	U
				CV	SD			
1	GLUCOSE III GOD-POD III	1841	175.58	5.13	9.00	176 mg/dL	0.05	0.42
2	CREATININE III Jaffes Kinetic - Alkaline Picrate III	1852	1.76	6.77	0.12	1.89 mg/dL	1.09	0.01
3	CHOLESTROL III CHOD-PAP III	1764	113.80	4.07	4.64	106 mg/dL	-1.68	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.00	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Pethampalayam
THE MEDICAL OFFICER, PERUNTHURAI

S. Senthirapathy
S. Senthirapathy
THE MEDICAL OFFICER, PERUNTHURAI