

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Adhanur									
Lab No	12488								<a href="#">Details About Robust Analysis</a>	
Year	2023								<a href="#">Detail About Monthly Summary</a>	
Month	June								<a href="#">Detail about SDI</a>	
Constituent Group	Chemistry III									
<a href="#">All Method Result</a>			<a href="#">Print</a>							
Result Entered Date : 20/06/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	268 mg/dL	-1.26	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	110 mg/dL	-0.74	0.24
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)

