

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Adhanur									
Lab No	12488								<a href="#">Details About Robust Analysis</a>	
Year	2023								<a href="#">Detail About Monthly Summary</a>	
Month	May								<a href="#">Detail about SDI</a>	
Constituent Group	Chemistry III									
<b>All Method Result</b>				<b>Print</b>						
Result Entered Date : 20/05/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	180 mg/dL	-0.09	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	121 mg/dL	1.10	0.27
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)

