



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 15.09.2023 Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE , NAMMIYAMPATTU, JAMUNAMARATHUR, THIRUVANNAMALAI.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMO SCIENTIFIC
ii Model	FINEPITTE
iii Value	10 µl -100 µl
iv Serial No	QW10090

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.03	100.05
2	10.04	50.05	100.05
3	10.00	50.02	100.04
4	10.02	50.08	100.08
5	10.00	49.99	100.04
6	10.02	50.07	100.02

For Heart Beat Medical Systems,


(Authorized signatory)

