

## VIEW LAB MONTHLY SUMMARY


Lab Name	Govt. PHC, Keeraipatti	Details About Robust Analysis	
Lab No	12109	Detail About Monthly Summary	
Month	July	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		
Click on the analyte to view Graphical Data		<a href="#">All Analyser Result</a>	<a href="#">Print</a>
Date of Result Entered : 19/07/2023			
Date of Report Published : 07/08/2023			

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	105 mg/dL	-0.72	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	109 mg/dL	-0.86	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99,	Accept with caution Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
 கருண்மணி  
 அரசு ஆரம்ப கல்விநிலையம்  
 கீரைப்பட்டி - 636903.  
 திருச்சி