



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: <b>Patliputra Nursing Home</b> Lab Name -7, Samepur Patliputra Clinical Lab Delhi 110042	Service request No. & date	P/01/09/06/23&09/06/2023
		ULR no.	CC223923000003062 F
		Cert. No.	AT/23000003062
		Date of Receipt of DUC	09/06/2023
		Date of calibration	09/06/2023
		Date of issue	10/06/2023
		Suggested due date	08/06/2024

Instrument Details			
Instrument name	Micro Pipette	Location	----
Make / Model No.	Dragon Lab	Accuracy	----
Range / Size	100 to 1000 µl	Visual Inspection	OK
Least Count	5 µl	Calibration Performed at	
I.D. No. / Sr. no.	--- / YE228AX0496796		

Detail of reference standards & Major equipments used			
Equipment Name	Digital Weighing Balance	A set of weight box	
Make	AND	WEIGHTRONICS	
Model / SR No.	GH-252/ 78002	----	
Certificate No.	TYCON/WB/02/23/327	TYCON/WB/12/274	
Calibration Validity	19/02/2024	25/12/2023	
Calibration by	TYCON ENGINEERING	Tycon Engineering	

Environmental Condition	Temperature	(23± 3) °C	Calibration Reference	IS:4162 (1985)
	Relative Humidity	(50±20) %	Work Instruction	AT/WI-20M

Calibration Results			
Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	100	100.25	±0.46 µl
02.	300	300.56	
03.	500	499.84	
04.	800	798.98	
05.	1000	999.91	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By  
(Calibration Engg./TM)  
(PRASHANT)  
Form No.- QF-47

Approved By  
(QM/TM)  
(NEERAJ TYAGI)  
Page No. 1 of 1



---End of Report---