

16/23, 12:57 PM

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UGPHC, Thiruvegampathu	Details About Robust Analysis
Lab No	7178	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle	Apparatus	No of Participants	DV	Participants	Year Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	82 mg/dL	-1.39	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.39 mg/dL	0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	115 mg/dL	0.70	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Acceptable, Warning Signal.
Beyond ±3.0	Unacceptable Performance, Action Signal.

Links

Modify Address	Change Mail	View Lab Results	View Lab Monthly Summary
Yearly Summary	Change Pa		Sign-out

(Handwritten Signature)
 உதவி இயக்குநர்
 சுகாதார அமைச்சு
 திருச்சி-630 408

5/31/23, 1:28 PM



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. UGPHC, Thiruvegampathu Lab No 7178
 Constituent Group Chemistry III Date of Result Entered : 19/04/2023
 PT item Lyophilized human serum based Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	120 mg/dL	0.08	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110 mg/dL	-0.28	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UGPHC, Thiruvegampathu
 The Medical Officer, Thiruvegampathu, Sivagangai District
 Thiruvegampathu
 TAMILNADU630408

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

மருத்துவ அறிவுரை:
 இந்த அறிவுரை மருத்துவ நோயாளிகளுக்கு
 திருவேகம்புத்தி-630 408
 சிவகங்கை மாவட்டம்

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UGPHC, Thiruvogampathu	Details About Robust Analysis
Lab No	7178	Detail About Monthly Summary
Month	May	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	151 mg/dL	-3.18	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.6 mg/dL	-1.21	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	122 mg/dL	1.27	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

H. 102

[Handwritten Signature]

மருத்துவ சேவையளி
 அரசு சார்பில் கருநாடகம் திருவள்ளூர்
 தொடுவெல்லூர்-630 408
 மருத்துவமனை

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UGPHC, Thiruvegampathu
 Lab No: 7178
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	283 mg/dL	-0.24	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	117 mg/dL	0.64	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

M. D. ...
Dis. ...

[Signature]

சென்னை மருத்துவ கல்வி நிறுவனம்
 மருத்துவ கல்வி நிறுவனம்
 தொலைபேசி-630 408
 சென்னை, இந்தியா

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UGPHC, Thiruvegampathu
 Lab No: 7178
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

- Details About Robust Analysis
- Detail About Monthly Summary
- Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 17/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	120 mg/dL	1.50	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Shrini
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 திருவேகம்பத்தி-630 473
 சிவகங்கை மாவட்டம்

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UGPHC, Thiruvegampathu
 Lab No: 7178
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/08/2023
 Date of Report Published : 01/09/2023

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	133 mg/dL	-4.73	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2.1 mg/dL	2.86	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	108 mg/dL	-1.25	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

(Signature)
 மருத்துவ அலுவலர்
 தலை அமைப்பு சுகாதார நிலையம்
 திருச்செங்கற்பட்டி-630 478
 சிவசங்கர மாவட்டம்