

VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. PHC, Melavilvarayanallur	<b>Details About Robust Analysis</b>
<b>Lab No</b>	12921	<b>Detail About Monthly Summary</b>
<b>Month</b>	March	<b>Detail about SDI</b>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	93.4 mg/dL	0.28	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	114.2 mg/dL	0.56	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

**மருத்துவ அலுவலர்**  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 மேல்வில்வராயநல்லூர்  
 தி.மலை மாவட்டம்-606006  
 GroupId=9&y...

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2\_new.aspx?resDetId=263&noSamples=1&month=3&year=2023&group...