

Lab Name: Govt. PHC, Anandai

Lab No: 12948

Month: August

Year: 2023

Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

[Click on the analyte to view Graphical Data](#)

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 172 mg/dL | -0.40 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.6 mg/dL | -1.34 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 109 mg/dL | -1.04 | 0.22 |

SDI Range

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within +1.01 to +2.00 | Good. |
| Within +2.01 to +2.99 | Accept with caution. Warning Signal. |
| Beyond +3.0 | Unacceptable performance. Action Signal. |

H. G. V.

Medical Officer.

**Govt. Primary Health Centre
Anandai-606 902.**

Aruvannamalai District

| | | | |
|----------------|-----------------|--------------------------|----------|
| Links | | | |
| Modify Address | Modify Methods | View Lab Monthly Summary | Sign-out |
| Yearly Summary | Change Password | | |